		<b>ENERGY CONTR</b>	ROL PRO	DCEDURE	
Plant:				Developed By:	
Department/Proce	55:			Reviewed By:	
Equipment Name:		Thermotron 20	Origin Date:		1/31/2019
Asset Number:				Revision Date:	2,02,202
Procedure Purpos	e and Compl	lance			
		establishes the minimum requireme	nts for the locko	out of energy isolat	ing devices whenever maintenance
W		n machines or equipment as outline			
authorized employe result in injury to pe	es are require ersonnel or dan	quired to comply with the restriction d to perform the lockout in accordar nage to equipment and may result in he following tasks associated with th	nce with this pro n disciplinary act	cedure. Failure to	follow this lockout procedure may
	1 Maintenance	and Servicing	_ 3_		
	2	140	_ 4_		
2		# Locks Needed		Special I	nstruction
_		for Lockout	Disconnect Nitr	ogen at the source o	n the wall.
		Cautionary	Stateme	ent	
All employees and	contractors wo	rking under this lockout procedure			k to each isolation point.
Lockout Sequence					
STEP 1	Notify all affected employees that the equipment must be shut down and locked out.				
STEP 2	Authorized e	mployee shall understand the hazar	ds of the energy	and shall know the	e methods to control the energy.
STEP 3	Shut equipm	ent down by the normal stopping pr	ocedure		
STEP 4	De-activate the energy isolating device(s) so that the machine or equipment is isolated from the energy source(s) identified below.				
Energy Source Magnitude	Isolation Point ID	Energy Isolating Device & Isolation Method	Lockout Device	Stored Energy?	Zero Energy Verification
Electrical 480 VAC	760	Place disconnect in off position and apply lock.	Lock	No	Actuate controls
Gas 51 - 100 PSI	No Tag	Turn valve to closed position and apply lockout device and lock.	Gate Valve Device	No	Check for Pressure
STEP 5	Lock out the energy isolating device(s) with assigned individual lock(s) or process locks.				
STEP 6	Stored or residual energy must be dissipated or restrained as shown below.				
Energy Source	Method of Control or Dissipation. Equipment Need				
STEP 7	Verify the isolation of the equipment by operating control(s) or by testing to make certain the equipment will not operate by following the Zero Energy Verification outlined in section 4. Please note that electrical work or access to electrical conductor requires zero energy verification with a properly rated meter.				

49202, Thermotron 20 Page 1 of 2

The machine or equipment is now locked out.

STEP 8

## **ENERGY CONTROL PROCEDURE**

Plant:

Developed By:

Reviewed By:

Equipment Name:

Thermotron 20

Asset Number:

Developed By:

Reviewed By:

1/31/2019

Revision Date:

Equipment Photo: 49202, Thermotron 20



## **Isolation Point and Controls Identification**

Description: LOTO# 760, 480 VAC Disconnect Description: LOTO# No Tag, Gate Valve
Location: Location: Wall





## **Return to Service**

Step 1 Verify equipment and area is clear of tools, workers, equipment, materials, and debris.

Step 2 Verify controls are in neutral.

**Step 3** Reposition any safety devices, guards, interlocks.

**Step 4** Warn workers to stay clear of area.

Step 5 Remove all locks and tags from energy control points.

Step 6 Verify affected areas are clear of personnel.

Step 7 Re-energize the machine or equipment.

49202, Thermotron 20 Page 2 of 2